

# EEG Electrode Set

## BE-403A

### General

The BE-403A EEG electrode set is designed for EEG measurement and EEG examination. This EEG electrode set can be used with a Nihon Kohden EEG-1200A/J/K, EEG-1250, EEG-9100A/G/J/K or EEG-9200A/G/J/K electroencephalograph.

### Safety Information

**WARNING** A warning alerts the user to possible injury or death associated with the use or misuse of the instrument.

**CAUTION** A caution alerts the user to possible injury or problems with the instrument associated with its use or misuse such as instrument malfunction, instrument failure, damage to the instrument, or damage to other property.

Pay attention to all safety information in the Operator's Manual or Installation Guide.

### WARNING

- Do not use this electrode during MRI examination because it may cause skin burn on the electrode placement area.
- Do not use the ESU with the electrode. If the ESU is used with the electrode, current from the ESU flows into the electrodes and causes electrical burn where the electrodes are attached.
- Before defibrillation, remove from the patient all electrodes connected to the shield junction cord. Otherwise, the operator may receive electrical shock.

### CAUTION

- Do not attach the electrode on a wound or inflamed skin.
- Do not use this electrode for electric stimulation.
- Do not store this electrode in an atmosphere with sulfur. It will deteriorate the electrode.

### CAUTION

Do not clean the electrode with hot water or chloric bleaching agent.

### CAUTION

Dispose of Nihon Kohden products according to your local laws and your facility's guidelines for waste disposal. Otherwise, it may affect the environment. If there is a possibility that the product may have been contaminated with infection, dispose of it as medical waste according to your local laws and your facility's guidelines for medical waste. Otherwise, it may cause infection.

This Safety and Performance Information is an extract from the general and safety information sections of the most recent edition of Operator's Manual or Installation Guide. Therefore, the contents of your Operator's Manual or Installation Guide may differ from those of this Safety and Performance Information. For detailed operating procedures, follow the instructions of your Operator's Manual or Installation Guide.

 Manufacturer  
**NIHON KOHDEN CORPORATION**  
 1-31-4 Nishiochiai, Shinjuku-ku,  
 Tokyo 161-8560, Japan  
 Phone +81 3-5996-8041  
<https://www.nihonkohden.com/>

 European Representative  
**NIHON KOHDEN EUROPE GmbH**  
 Raiffeisenstrasse 10, D-61191 Rosbach, Germany  
 Phone +49 6003-827-0 Fax +49 6003-827-599