

Electrocardiograph

ECG-3150

General

The Shanghai Kohden ECG-3150 electrocardiograph is intended for medical use to process the electrical signals produced by the heart, which are acquired through two or more electrodes, and to display waveforms and/or prepare a record of these electrical signals. This device is a portable ECG acquisition terminal which measures up to 12-lead ECG waveforms. The intended use is only for diagnosis, not for monitoring of vital physiological parameters.

ECG-3150 is intended to be used by qualified medical personnel within a medical facility, such as hospital or clinic and children and adults can be diagnosed with this electrocardiograph.

Safety Information

- ⚠ DANGER:** A danger alerts the user to a hazardous situation which causes death or serious injury.
- ⚠ WARNING:** A warning alerts the user to possible injury or death associated with the use or misuse of the instrument.
- ⚠ CAUTION:** A caution alerts the user to possible injury or problems with the instrument associated with its use or misuse such as instrument malfunction, instrument failure, damage to the instrument, or damage to other property.

Pay attention to all safety information in the Operator's Manual or Installation Guide.

⚠ WARNING

Never use the electrocardiograph in the presence of any flammable anesthetic gas or high concentration oxygen atmosphere. Failure to follow this warning may cause explosion or fire.

⚠ WARNING

Never use the electrocardiograph in a hyperbaric oxygen chamber. Failure to follow this warning may cause explosion or fire.

⚠ WARNING

When the electrocardiograph is used with an electrosurgical unit (ESU), firmly attach the entire area of the ESU return plate. Otherwise, the current from the ESU flows into the electrodes of the electrocardiograph, causing electrical burn where the electrodes are attached. For details, refer to the ESU manual.

⚠ WARNING

When performing MRI test, remove all electrodes and transducers from the patient which are connected to this device. Failure to follow this warning may cause skin burn on the patient. For details, refer to the MRI manual.

⚠ WARNING

Before defibrillation, remove everything including electrodes and patches from the patient's chest. If the defibrillator paddle contacts an object on the patient's chest, the discharged energy may be insufficient and cause skin burn.

⚠ WARNING

Before defibrillation, all persons must keep away from the bed and must not touch the patient or any equipment or cable connected to the patient. Failure to follow this warning may cause electrical shock or injury.

⚠ WARNING

Use only the specified patient cables. Failure to follow this warning may cause skin burn where the electrode is attached and damage the electrocardiograph due to discharge energy when defibrillation is performed.

⚠ WARNING

No modification of this equipment is allowed.

⚠ WARNING

Transport only under conditions described in instructions for use or marked on ME EQUIPMENT with an indication of RESIDUAL RISK if ME EQUIPMENT or its parts overbalances.

⚠ WARNING

This device is used for diagnosis only.
Do NOT use for monitoring in ICU/CCU or emergency room.

⚠ WARNING

Only connect Shanghai Kohden specified device to the electrocardiograph and follow the specified procedure. Failure to follow this warning may result in electrical shock or injury to the patient and operator, and cause fire or device malfunction.

⚠ WARNING

Install non-medical devices which are connected to the electrocardiograph outside the patient environment (IEC 60601-1:2012). If they are installed inside the patient environment, the patient or operator may receive electrical shock or injury.

⚠ WARNING

Additional equipment connected to the medical electrical equipment must comply with the respective IEC or ISO standards (e.g. IEC 60950 for data processing equipment). Furthermore all configurations shall comply with the requirements for medical electrical systems (see IEC 60601-1). Anybody connecting additional equipment to medical electrical equipment configures a medical system and is therefore responsible that the system complies with the requirements for medical electrical systems. Attention is drawn to the fact that local laws take priority over the above mentioned requirements. If in doubt, consult your local representative or the technical service department.

⚠ WARNING

When the electrocardiograph is connected to an external device, ground all devices to the same one-point ground and use the specified isolation transformer for the external device even when the electrocardiograph is operated on battery pack. Otherwise, the patient and operator may receive electrical shock from leakage current from the external device.

⚠ WARNING

When several medical devices are used together, ground all devices to the same one-point ground. Any potential difference between devices may cause electrical shock to the patient and operator.

⚠ WARNING

Do not connect the USB hub power cord to a wall power socket. Failure to follow this warning may cause electrical shock to the patient and operator.

⚠ WARNING

Use only the provided power cord. Otherwise, the patient and operator may receive electrical shock or injury.

⚠ WARNING

When the external device does not comply with the IEC 60601-1, use a local purchase medical isolation transformer unit between the external device and the wall power socket.

⚠ WARNING

Do not use an additional multiple portable socket outlet or extension cord. The impedance of the protective earth increases and the patient and operator may receive electrical shock or injury.

⚠ WARNING

Do not put the multiple portable socket outlet on a floor. If the multiple portable socket outlet gets dusty, damaged or contacts liquid, it may cause device malfunction and electrical shock to the patient or operator.

⚠ WARNING

Install the laser printer which is connected to the electrocardiograph outside the patient environment (IEC 60601-1). If the laser printer is installed inside the patient environment, the patient or operator may receive electrical shock or injury. For installation, contact your Nihon Kohden representative.

⚠ WARNING

To avoid the risk of electric shock, this equipment must only be connected to a supply mains with protective earth.

⚠ WARNING

Additional multiple socket-outlets or extension cords must not be connected to the system.

⚠ WARNING

The multiple socket-outlet, when provided as a separate item must not be placed on the floor. This may cause electric shock.

⚠ WARNING

Connect only items that have been specified as part of the system or specified as being compatible with the system.

⚠ WARNING

To use non-medical equipment that requires an isolation transformer, only connect it to a specified multiple portable socket outlet that has an isolation transformer. Do not connect the non-medical equipment to a wall power socket or to a multiple portable socket that does not have an isolation transformer. Do not connect unspecified electrical devices to a multiple portable socket outlet in the system. If unspecified electrical devices are connected, leakage current increases and the patient and operator may receive electrical shock or injury.

⚠ WARNING

Follow IEC 60601-1 for connecting the electrocardiograph to other instruments.

⚠ WARNING

- Never short-circuit the + and – terminals on the battery pack. It may cause overheating and fire.
- Keep the battery pack away from fire. It may explode.
- Do not damage, disassemble, drop or give impact to the battery pack.
- Never use the battery pack on unspecified devices.
- Never charge the battery pack on unspecified devices.
- Never install the battery pack with the wrong polarity.
- Keep the battery pack away from patients and out of the reach of children.

⚠ WARNING

- Keep the battery pack away from fire, and never heat the battery pack, or it may result in leakage and explosion of the battery pack.
- Do not handle or store the battery pack with metals like necklace or hair clipper, because this has the risk of leakage and explosion due to short circuit of the battery pack.
- Do not use a battery pack which is damaged or dropped. There is a gas release valve inside the battery pack, and if the valve is damaged due to dropping, etc. the gas won't be released during operation, and this will induce explosion.

⚠ WARNING

If the battery pack is damaged and the substance inside the battery pack contacts the eyes or skin, wash immediately and thoroughly with water and see a physician. Never rub your eyes, because you may lose your eyesight.

⚠ WARNING

- Do not immerse the battery pack in water. The battery pack may overheat and rust and the substance inside the battery pack may leak.
- Do not leave the battery pack unused for more than about one year. The battery pack may leak.

⚠ WARNING

Only use the SB-301DC battery pack.

⚠ WARNING

Never connect the battery pack to an AC jack or ignition socket of an automobile. Failure to follow this warning may result in leakage and explosion of battery pack.

⚠ WARNING

It's strictly forbidden to disassemble, reconstruct, or damage the battery pack, or weld the shell directly. Failure to follow this warning may result in leakage and explosion.

⚠ WARNING

- Avoid strong mechanical impact to the battery pack, may cause leakage and explosion of battery pack.
- Do not use a battery pack which is damaged, polluted or leaking. Failure to follow this warning may result in explosion.

⚠ WARNING

Check the direction of the connector when connecting the battery pack cable. Don't force the battery pack cable into the battery pack connector if abnormality is felt.

⚠ WARNING

Do not charge the battery pack with any equipment other than the electrocardiograph. Otherwise, abnormal current will cause the leakage and explosion of battery pack.

⚠ WARNING

Remove the battery pack from the electrocardiograph when it will not be used for more than 1 month. Otherwise the battery pack may leak and rust.

⚠ WARNING

Additional equipment connected to the medical electrical equipment must comply with the respective IEC or ISO standards (e.g. IEC 60950 for data processing equipment). Furthermore all configurations shall comply with the requirements for medical electrical systems (see IEC 60601-1). Anybody connecting additional equipment to medical electrical equipment configures a medical system and is therefore responsible that the system complies with the requirements for medical electrical systems. Attention is drawn to the fact that local laws take priority over the above mentioned requirements. If in doubt, consult your local representative or the technical service department.

⚠ WARNING

Enter the patient information correctly. Otherwise, the ECG data may be lost or mixed up with another patient's ECG data. When managing test data that is only in electronic file, make sure to enter the patient information. After sending the DICOM data to an external system (such as PACS, etc.), make sure the ECG data is received and displayed correctly and the displayed patient information are consistent with the entered patient information.

⚠ WARNING

Connect the electrocardiograph to a network as specified. Otherwise the patient and operator may receive electrical shock or injury. To connect the network, contact your Nihon Kohden representative.

⚠ WARNING

Do not connect the tip of the electrode lead to anything other than an electrode. Connect all the tips to the corresponding electrodes and attach them to the patient. If unconnected tips or electrodes contact an item which is not equipotentially grounded, the patient may receive electrical shock. Also, noise may superimpose on the ECG waveforms and measurement data may be incorrect.

⚠ WARNING

Only use the BJ-961D, BJ-962D, BJ-901D, BJ-902D, BJ-903D or BA-901D patient cables when using with a defibrillator. When the specified patient cable is connected, the electrocardiograph is type CF defibrillation-proof compliance. Failure to follow this warning may cause skin burn where the electrode is attached and damage the electrocardiograph due to discharge energy when defibrillation is performed.

⚠ WARNING

Portable RF communications equipment (including peripherals such as antenna cables and external antennas) should be used no closer than 30 cm (12 inches) to any part of the ECG-3150, including cables specified by the manufacturer. Otherwise, degradation of the performance of this equipment could result.

⚠ WARNING

Do not touch the patient while touching any metal part of non-medical electrical equipment or an exposed part when a connector or cover is removed. Failure to follow this warning may cause electrical shock or injury to the patient.

⚠ WARNING

Do not use a damaged network cable. The patient or operator may receive electrical shock when the damaged part is touched.

⚠ WARNING

Do not diagnose a patient based on DICOM or PDF data acquired by the electrocardiograph. Otherwise, it may cause incorrect judgment.

⚠ WARNING

The operator should not touch patient and the input/output interface of the instrument simultaneously. This may cause electric shock.

⚠ WARNING

- Replacement of a component could result in an unacceptable risk, cause electric shock or discard the device.
- Only promise that the service personnel can replace the component.

⚠ CAUTION

Before connecting or disconnecting devices, make sure that each device is turned off and the power cord is disconnected from the power socket. Otherwise, the patient or operator may receive electrical shock or injury.

⚠ CAUTION

Do not install the electrocardiograph where it will be exposed to water or chemical solutions. Avoid direct sprinkling, spray or moist air from a nebulizer or humidifier. These cause malfunction and shorten the life of the electrocardiograph.

⚠ CAUTION

Do not place electrocardiograph upright. Failure to follow this instruction may cause the electrocardiograph to fall over.

⚠ CAUTION

When installing the electrocardiograph, take care not to pinch your finger under the electrocardiograph.

⚠ CAUTION

The electrocardiograph should not share the power line with any device which consumes a lot of power, such as an X-ray device, because it may cause artifact.

⚠ CAUTION

Avoid locations where the electrocardiograph may receive strong electromagnetic interference such as radio or TV stations, mobile phones or mobile two-way radios.

⚠ CAUTION

- Do not install the electrocardiograph near a power line, dynamo or motor which has electromagnetic induction.
- Do not install the electrocardiograph near an electrosurgical unit or RF therapeutic equipment.
- Select a room without excessive noise, vibration, sunlight, high humidity or water splashes.
- Make sure that there is no influence from a mobile phone.

⚠ CAUTION

Make sure that there is enough space between the electrocardiograph and the wall for adequate ventilation. Leave more than 5 cm of space between the wall and electrocardiograph so that the operating temperature does not exceed 40 °C (104 °F). Otherwise the internal temperature of the electrocardiograph rises, which leads to inaccurate operation and shortens the electrocardiograph life.

⚠ CAUTION

Always install the battery pack even when the electrocardiograph operates on AC power. Otherwise sudden power off occurs when an electrode is detached during recording.

⚠ CAUTION

Do not bend or pull the connector of the battery pack with strong force. This may damage the battery pack cable or battery pack cable connector.

⚠ CAUTION

If the operation time of a fully charged battery pack is less than 30 minutes, the battery pack may have become deteriorated. Do not use a deteriorated battery pack. Charging a deteriorated battery pack may cause the electrocardiograph to fail to operate in both battery pack and AC power.

⚠ CAUTION

Replace the battery pack with a new one every year to assure maximum performance from the electrocardiograph.

⚠ CAUTION

- Charge the battery pack at ambient temperatures of 5 °C to 40 °C (41 °F to 104 °F). If the battery pack is charged below 5 °C or over 40 °C, the battery pack may leak or heat up. This may damage the battery pack.
- Store the battery pack at temperatures:
 - 20 °C to +65 °C (Less than 1 week)
 - 20 °C to +55 °C (Less than 1 month)
 - 20 °C to +45 °C (Less than 6 months)
 - 20 °C to +35 °C (Less than 1 year) in a dry place to prevent rusting and leakage.

⚠ CAUTION

Before inserting or removing battery pack, make sure that the electrocardiograph is turned off and the power cord is disconnected from the power socket. Otherwise, the operator may receive electrical shock.

⚠ CAUTION

Do not expose the battery pack to direct sunlight or leave in a high temperature place. The lifetime of the battery pack may be shortened, the performance of the battery pack may be degraded and the battery pack may leak.

⚠ CAUTION

If the ambient temperature is over 30 °C (86 °F), charge it every three months because it self-discharges quickly.

⚠ CAUTION

When the battery pack is stored more than 1 month, charge it before use because the battery pack may have self-discharged.

⚠ CAUTION

The battery pack must be replaced by qualified service personnel.

⚠ CAUTION

The SB-301DC battery pack is Ni-MH (Nickel-Metal Hydride). Before disposing of the battery pack, check with your local solid waste officials for details in your area for recycling options or proper disposal. The battery pack is recyclable. At the end of its useful life, under various state and local laws, it may be illegal to dispose of this battery pack into the municipal waste stream.

⚠ CAUTION

Do not leave accessories and consumables such as electrodes or electrolyte cream (Cardio Cream) near the patient or in reach of children. If they are swallowed, consult a physician immediately.

⚠ CAUTION

Only use Shanghai Kohden specified parts and accessories to assure maximum performance from your electrocardiograph. Using unspecified recording paper or electrodes may cause incorrect ECG recording and accelerate deterioration of the electrocardiograph.

⚠ CAUTION

Only use a Shanghai Kohden specified USB flash disk. Otherwise, the USB flash disk may not be available.

⚠ CAUTION

Do not leave the USB flash disk near patients or in reach of children.

⚠ CAUTION

Only use the Shanghai Kohden specified 63 mm recording paper. When a narrower recording paper is used and the paper detection circuit cannot detect it, the thermal head may be activated. The thermal head may be damaged where there is no recording paper and the rubber paper feed roller may get burned.

⚠ CAUTION

Prolonged storage under conditions such as high temperatures, excessive humidity, direct sunlight, and fluorescent light causes fading of the colored surface. Avoid high temperatures exceeding 50 °C (122 °F) and excessive humidity. Store the recording paper in a cool, dry and dark place if possible.

⚠ CAUTION

- If unspecified recording paper is used, it might not print properly or send multiple pages.
- Follow the procedure to set the recording paper. Otherwise, the recording paper might not feed normally or the feeding motor may stop operation.
- When pasting the recording paper, use paste with a starch base, PVA base, arabic gum base paste, CMC base synthetic paste, or paper bond.

⚠ CAUTION

- Do not apply high pressure to the recording paper. Rubbing or scratching the surface with a hard object discolors it.
- Do not use recording paper which is contaminated with ECG electrolyte cream (cardioCream) or saline solution because the recording paper will fade by the chloride and the contaminated thermal head may cause missing dots. If the thermal head gets contaminated, clean it with the thermal head cleaner pen.
- Do not set the recording paper immediately after recording because the thermal head is still hot.

⚠ CAUTION

Organic solvents and plastics may develop color on the white surface or cause the recording paper to fade. Examples of color-developing materials containing organic solvents: adhesives, paste, felt-tip pens, and semi-dry diazo paper. Examples of color-fading materials made of soft vinyl chloride: document cases, albums, and desk pads including. Examples of color-fading materials containing plastics: erasers, adhesive tapes, and some fluorescent pens.

⚠ CAUTION

- Clean the thermal head with the thermal head cleaner after every 100 m printing. A dirty thermal head degrades printing quality.
- Do not touch the thermal head with your fingers. If accidentally touched, clean the thermal head with the thermal head cleaner pen.

⚠ CAUTION

- Never disassemble or modify the USB flash disk.
- Do not bend or give a strong impact to the USB flash disk.
- Do not handle the USB flash disk while smoking or eating. Do not get the USB flash disk wet.
- Do not touch any metal part of the USB flash disk.
- Do not peel off the label or sticker on the USB flash disk or put another label or sticker on the USB flash disk.
- Only use the USB flash disk in the specified device.
- Do not scratch the insertion part of the USB flash disk. The electrocardiograph might not operate if dust gets into the insertion part.

⚠ CAUTION

To ensure the cybersecurity of the electrocardiograph, implement the following security measures in the network environment to which the product is connected.

1. All communication (incoming and outgoing) between the electrocardiograph and the local area network (HIS, etc.) is subject to packet filtering by a firewall or router.
2. When the electrocardiograph is connected to the local area network (HIS, etc.), all communication with the internet or other external networks is restricted to essential transmissions under the supervision of the appropriate personnel with responsibility for information security of medical equipment.

⚠ CAUTION

- Put the USB flash disk in the case when the USB flash disk is removed from the electrocardiograph. Store it where patients cannot accidentally swallow the USB flash disk.
- Do not expose the USB flash disk to direct sunlight or leave in a place of high temperature and high humidity.
- Operating environment
Temperature: -25 °C to +85 °C (-13 °F to +185 °F)
Humidity: -25 % RH to 95 % RH (noncondensing)
- Storage environment
Temperature: -40 °C to +85 °C (-40 °F to +185 °F)
Humidity: -25 % RH to 95 % RH (noncondensing)
- Do not store the USB flash disk in a corrosive gas environment.

⚠ CAUTION

The network must be managed by the network administrator. Only the network administrator can change the network settings on the ECG-3150 electrocardiograph and connect the electrocardiograph to the network. Incorrect settings or connection may cause failure of the network system and device.

⚠ CAUTION

Select an installation location where patient information cannot be seen or accessed by unauthorized personnel.

⚠ CAUTION

Use the electrocardiograph in a securely managed environment.

⚠ CAUTION

Personal information stored on the electrocardiograph, or the PC, is vulnerable to unauthorized access. Follow the provisions of the user agreement for the electrocardiograph related to information security.

⚠ CAUTION

Information can be browsed through web client only when a medical institution is compliant with ISMS (Information Security Management System). Accessing the information from a global network environment is not allowed.

⚠ CAUTION

When sending/receiving patient test data or patient information, the communication should be carried out through the security network environment inside the medical institution.

⚠ CAUTION

Only use the KD-105D/105E, KD-108D/108E cart for this electrocardiograph. If another cart is used, it may tip over or the electrocardiograph may fall off.

⚠ CAUTION

- Do not move the cart with the casters locked. The cart may tip over.
- When moving the cart, hold the handle only.

⚠ CAUTION

When moving the cart with patient cable hanger mounted, fold the upper and lower arms together, rotate the arms so that they point toward the handle and do not protrude beyond the edges of the cart.

⚠ CAUTION

To prevent the cart from tipping over or the electrocardiograph falling off the cart:

- Do not put or hook anything on the handle.
- Do not ride on the cart. Do not lean on the handle or put your weight on the cart.
- Periodically check that the casters, handle and frame are firmly attached and not wobbly.
- Always lock the casters so that the cart does not move accidentally.

⚠ CAUTION

Do not use the output signal from the output connector of the electrocardiograph for a synchronization signal on a defibrillator. There is a time delay between the input signal and output signal. When using the output signal from the electrocardiograph for the synchronization signal on other device, always consider this time delay.

⚠ CAUTION

The paper cutter in the recording paper container is sharp. When setting the recording paper, be careful not to cut your finger. Set the recording paper as specified.

⚠ CAUTION

Enter the patient information correctly. Otherwise, the ECG data may be lost or mixed up with another patient's ECG data.

⚠ CAUTION

The filter, auto gain and layout of the DICOM or PDF data which is output from the electrocardiograph are not entirely consistent with those on the recording paper.

⚠ CAUTION

Changing the number of digits in ID will delete all ECG data in the internal memory. Back up the necessary data before changing the number of digits in ID.

⚠ CAUTION

Do not remove the USB flash disk or turn the power off while formatting or data is being written to or read from the USB flash disk. This may damage the USB flash disk or saved data.

⚠ CAUTION

Only format or write data to the USB flash disk in the specified device. Otherwise, the USB flash disk may become unusable.

⚠ CAUTION

Do not use new and old electrodes together, or re-usable and disposable electrodes together (such as clip electrodes, suction electrodes and disposable electrodes). If different types are used together, it may cause a high polarization voltage and prevent displaying and recording of the ECG waveforms.

⚠ CAUTION

- Format the USB flash disk before the first use.
- Format the USB flash disk only in an ECG-3150 electrocardiograph. The USB flash disk formatted in other ECG models cannot be used in this electrocardiograph.

⚠ CAUTION

- Insert only a Shanghai Kohden specified USB flash disk in the USB flash disk connector.
- Insert or remove the USB flash disk straight into or out of the USB flash disk connector.

⚠ CAUTION

Make sure that the electrocardiograph power is turned off, when inserting or removing a USB flash disk. Otherwise, the data in USB flash disk may get damaged and operation of the electrocardiograph may become unstable.

⚠ CAUTION

The parts of electrodes and other connectors should not contact other conductive parts including earth.

⚠ CAUTION

When using an ESU and defibrillator with the electrocardiograph, use silver chloride disposable electrodes.

⚠ CAUTION

During recording, if artifact superimposes on the ECG waveform, immediately stop the recording and remove the cause of the artifact. Refer to Troubleshooting in Section 9. Failure to follow this instruction causes overheating and damage to the electrocardiograph.

⚠ CAUTION

When baseline drift suppression is set to "Strong", distortion may occur in the ST segment with a decrease of heart rate when heart rate is less than 50. For bradycardia, set this item to "Weak". The "Weak" setting has characteristics conforming to the AHA recommendations.

⚠ CAUTION

Only use clean and undamaged electrodes with no scratches on the surface. Electrodes with a deteriorated surface may cause high contact impedance and distorted ECG waveforms.

⚠ CAUTION

Always replace all the electrodes with new ones at the same time.

⚠ CAUTION

- The automatic ECG interpretation is performed for acquired ECG waveforms only and does not reflect all conditions of the patient. The results of the analysis might not correspond to the judgement of a physician.
- Overall judgement must be performed by the physician, referring to the analysis result, clinical findings, and other examination results. After the physician's overall judgement, the analysis results should be signed or initialed by the physician.

⚠ CAUTION

Take care when interpreting the ECG recording because the EMG filter may cause distortion of P-waves and QRS-waves depending on the waveform shape.

⚠ CAUTION

Do not turn the power off during saving or sending ECG data file, otherwise the data may be lost.

⚠ CAUTION

When the patient sex and age is not entered, ECG analysis is performed with the patient as male and 35 years old.

⚠ CAUTION

The "CAUTION LABEL" contains important information for operating the electrocardiograph. Clean the label if it is dirty. If the label is damaged or difficult to read, contact local agents to replace it with a new one.

⚠ CAUTION

- Wipe the electrocardiograph thoroughly after disinfecting it with spray.
- Do not sterilize or disinfect the electrocardiograph with ultraviolet light or ozone.

⚠ CAUTION

Never sterilize the electrocardiograph because the materials may deform, crack or discolor.

⚠ CAUTION

After cleaning, make sure that the electrocardiograph is completely dried.

⚠ CAUTION

- Do not use a rough cloth.
- Do not use acidic, alkaline detergents, or alcohol other than ethanol or isopropyl.

⚠ CAUTION

- Do not use hypochlorous acid for disinfecting.
- Wipe the patient cable thoroughly after disinfecting it with spray.

⚠ CAUTION

- Do not clean the thermal head immediately after recording because the head is still hot.
- Only clean the head with the provided thermal head cleaner pen. Otherwise the thermal head may be damaged.

⚠ CAUTION

- Do not immerse the connector in liquid.
- Do not use volatile liquids such as thinner or benzene, because these will cause the materials to melt or crack.

⚠ CAUTION

Do not sterilize electrodes by boiling. It may damage the electrode.

⚠ CAUTION

Before maintenance, cleaning or disinfection, turn the electrocardiograph power off and disconnect the AC adapter. Failure to follow this instruction may result in electrical shock and electrocardiograph malfunction.

This Safety and Performance Information is an extract from the general and safety information sections of the most recent edition of Operator's Manual or Installation Guide. Therefore, the contents of your Operator's Manual or Installation Guide may differ from those of this Safety and Performance Information. For detailed operating procedures, follow the instructions of your Operator's Manual or Installation Guide.

 Manufacturer
SHANGHAI KOHDEN MEDICAL
ELECTRONIC INSTRUMENT CORP.
No.567 Huancheng Bei Road, Shanghai Comprehensive
Industrial Development Zone, Fengxian District,
Shanghai 201401, China
Phone +86 21-5743-6998
<https://www.nihonkohden.com/>

 European Representative
NIHON KOHDEN EUROPE GmbH
Raiffeisenstrasse 10, 61191 Rosbach, Germany
Phone +49 6003-827-0 Fax +49 6003-827-599

