

SpO₂ Connection Cord

JL-430U

General

The JL-430U SpO₂ connection cord connects a Nihon Kohden probe to a Nihon Kohden pulse oximeter with SpO₂ measurement.

Safety Information

CAUTION A caution alerts the user to possible injury or problems with the instrument associated with its use or misuse such as instrument malfunction, instrument failure, damage to the instrument, or damage to other property.

Pay attention to all safety information in the Operator's Manual or Installation Guide.

CAUTION

- Use the SpO₂ connection cord only with the specified instruments. If this cord is connected to an unspecified instrument, pulse signals may not be detected, measured values may be incorrect, or the patient may get skin burn.
- Use the SpO₂ connection cord with the specified attachment tape and probe.
- Avoid the location where will be exposed to water at the time of storage.
- Do not use a damaged, disassembled, altered or broken probe and SpO₂ connection cord.

CAUTION

Do not sterilize the SpO₂ connection cord. This may damage or deteriorate the SpO₂ connection cord.

CAUTION

Dispose of Nihon Kohden products according to your local laws and your facility's guidelines for waste disposal. Otherwise, it may affect the environment. If there is a possibility that the product may have been contaminated with infection, dispose of it as medical waste according to your local laws and your facility's guidelines for medical waste. Otherwise, it may cause infection.

This Safety and Performance Information is an extract from the general and safety information sections of the most recent edition of Operator's Manual or Installation Guide. Therefore, the contents of your Operator's Manual or Installation Guide may differ from those of this Safety and Performance Information. For detailed operating procedures, follow the instructions of your Operator's Manual or Installation Guide.



Manufacturer

NIHON KOHDEN CORPORATION
1-31-4 Nishiochiai, Shinjuku-ku,
Tokyo 161-8560, Japan
Phone +81 3-5996-8041
<https://www.nihonkohden.com/>



European Representative

NIHON KOHDEN EUROPE GmbH
Raiffeisenstrasse 10, D-61191 Rosbach, Germany
Phone +49 6003-827-0 Fax +49 6003-827-599