

Battery Pack

SB-301DC

General

The SB-301DC battery pack is a rechargeable battery for ECG-3150 electrocardiograph. Please read the ECG-3150 electrocardiograph operator's manual for handling and safety of the battery pack.

Safety Information

- ⚠ DANGER:** A danger alerts the user to a hazardous situation which causes death or serious injury.
- ⚠ WARNING:** A warning alerts the user to possible injury or death associated with the use or misuse of the instrument.
- ⚠ CAUTION:** A caution alerts the user to possible injury or problems with the instrument associated with its use or misuse such as instrument malfunction, instrument failure, damage to the instrument, or damage to other property.

Pay attention to all safety information in the Operator's Manual or Installation Guide.

⚠ WARNING

To avoid leakage, overheating, explosion, fire or skin burn:

- Do not give strong impact to or deform the battery.
- Do not use the battery in unspecified instruments.
- Do not disassemble or modify the battery.
- Do not use the battery pack at temperatures over 40 °C (104 °F).
- Do not put the battery pack into fire or heat it.

⚠ WARNING

If the battery pack is damaged and the substance inside the battery pack contacts the eyes, wash immediately and thoroughly with water and see a physician. Never rub your eyes. Otherwise, you may lose your eyesight.

⚠ WARNING

Failure to observe any of the following may cause battery malfunction, overheating, explosion, fire and skin burn:

- Do not immerse the battery pack in liquid or get it wet.
- Do not leave the battery near a heat source such as a stove.
- The battery pack could only be charged in specified instruments. Please charge the battery pack under specified environment. (Charging temperature cannot exceed 40 °C (104 °F)).
- Do not short-circuit the + and – terminals on the battery.
- Do not charge the battery in a high temperature place such as near a stove or in sun-heated cars.

⚠ CAUTION

Do not leave the battery pack in the reach of patients. Do not leave the battery pack in the direct sunlight or high-temperature places, such as in a sun-heated car. This may cause leakage of the battery pack, degrade its performance and make it unusable.

⚠ CAUTION

If the battery pack is damaged and the substance inside it contacts the skin or clothes, wash immediately and thoroughly with clear water. Otherwise, the skin may get irritated.

- NOTE**
- The battery pack lifetime may become shorter depending on the operating and storage environment.
 - When bringing the battery pack on an airplane, please consult the airline beforehand.
 - When transporting the battery pack by ship or by airplane, please consult the transport company beforehand.

⚠ CAUTION

Please dispose of Shanghai Kohden products according to your local laws and your facility's guidelines for waste disposal. Otherwise, it may affect the environment. If there is a possibility that the product may have been contaminated with infection, please dispose of it as medical waste according to your local laws and your facility's guidelines for medical waste. Otherwise, it may cause infection.

This Safety and Performance Information is an extract from the general and safety information sections of the most recent edition of Operator's Manual or Installation Guide. Therefore, the contents of your Operator's Manual or Installation Guide may differ from those of this Safety and Performance Information. For detailed operating procedures, follow the instructions of your Operator's Manual or Installation Guide.

 Manufacturer
 SHANGHAI KOHDEN MEDICAL
 ELECTRONIC INSTRUMENT CORP.
 No.567 Huancheng Bei Road, Shanghai Comprehensive
 Industrial Development Zone, Fengxian District,
 Shanghai 201401, China
 Phone +86 21-5743-6998 Fax +86 21-5743-6939
<https://www.nihonkohden.com/>

 European Representative
 NIHON KOHDEN EUROPE GmbH
 Raiffeisenstrasse 10, D-61191 Rosbach, Germany
 Phone +49 6003-827-0 Fax +49 6003-827-599