

**MEK-3DN HEMATOLOGY CONTROL 2.0mL**

**ASSAY SHEET**

<b>LOT</b>	B242N
	2024-05-05
<b>OPEN</b>	14DAYS

**MEK-1301,1302,1305 Barcode Entry**

The barcodes may be read in any order. If the barcode is difficult to scan, place this assay sheet on a flat surface, keep the barcode reader and barcode vertical, and scan within a distance of 2cm.

1



2



3



		RANGE	MEK -6400 -6410 -6420	MEK -6500 -6510	MEK -6450	MEK -6550	MEK -1301 -1302 -1305
<b>WBC</b>	10 <sup>3</sup> /μL	±0.8	7.9	7.9	7.9	7.9	7.75
<b>RBC</b>	10 <sup>6</sup> /μL	±0.20	4.66	4.66	4.66	4.66	4.66
<b>HGB</b>	g/dL	±0.5	13.9	14.1	13.9	14.1	13.95
<b>HCT</b>	%	±3.0	40.0	40.0	40.0	40.0	45.9
<b>MCV</b>	fL	±7.0	85.9	85.9	85.9	85.9	98.5
<b>MCH</b>	pg	±2.4	29.8	30.3	29.8	30.3	29.9
<b>MCHC</b>	g/dL	±3.0	34.8	35.3	34.8	35.3	30.4
<b>PLT</b>	10 <sup>3</sup> /μL	±50	278	278	278	278	281.8
<b>MPV</b>	fL	±2.0	5.8	8.2	5.8	8.2	8.4
<b>RDW-CV *</b>	%	±3.0	16.3	16.3	16.3	16.3	15.6
<b>LY%</b>	%	±15.0	33.8	33.8	33.8	33.8	34.64
<b>MO%</b>	%	±8.0	5.4	5.4	5.4	5.4	2.69
<b>GR%</b>	%	±15.0	60.8	60.8	60.8	60.8	62.67
<b>LY</b>	10 <sup>3</sup> /μL	±1.5	2.7	2.7	2.7	2.7	2.68
<b>MO</b>	10 <sup>3</sup> /μL	±0.8	0.4	0.4	0.4	0.4	0.21
<b>GR</b>	10 <sup>3</sup> /μL	±1.5	4.8	4.8	4.8	4.8	4.86

<b>MEK-1305</b>
<b>QC ESR</b>
<b>10±5 mm</b>

<b>Manual Registration</b>	
Lot	242
Level	Normal
Control	MEK-3D

\* RDW-CV is displayed as RDW on MEK-6400,6410,6420,6450.



<https://www.nihonkohden.com/spi/index.html>



**NIHON KOHDEN CORPORATION**  
1-31-4 NISHIOCHIAI, SHINJUKU-KU  
TOKYO 161-8560 JAPAN



0694-9016850